

Falsely decreased glucose

A case of falsely low blood glucose values due to a pre-analytical error.

A falsely normal OGTT result?

HOSP #		WARD	Vanguard Antenatal Clinic
CONSULTANT	Jody Rusch	DOB/AGE	32 y Female

Abnormal Result

Oral glucose tolerance test with the Sodium Fluoride (NaF) tubes registered by the lab 20 hours after being sampled.

Collection date: 07h22 05/03/2020

Received date: 18h44, 05/03/2020

Registered date: **03h26, 06/03/2020**

Please note that samples in our lab are being centrifuged after being registered.

Fasting Glucose	3.9	mmol/L
120 min. Glucose	4.4	mmol/L

Presenting Complaint

My thoughts were, that if the sample isn't centrifuged in a timely manner, metabolism would still happen, albeit at a slower rate. I also thought that metabolism (glycolysis) would

continue if left for a long period (>8 hours) uncentrifuged.

Would you argue the result as given below at "Laboratory Investigations" is reliable, given the following info?

1. Stability of glucose in whole blood in NaF tubes?
2. Could this be a false normal result?

The stability spreadsheet as summarized by our lab did not have the stability info for glucose in whole blood:

	Analyte	Synonym	After request	NSS Stability Limit (Room/20-25°C)(Serum)	NSS Stability Limit (Room/20-25°C)(Whole blood)	Stability Limit (Serum) 2-4°C (4-8°C if WHO)	Serum stability limit - 20°C	Serum Stability Limit Source	Comments
93	Glucose - NaF plasma		3 days	3 days		3 days		INSERT	

Figure 1: Stability spreadsheet as summarized by our lab did not have the stability info for glucose in whole blood.

History

Patient is most likely pregnant (being from an antenatal clinic) and this is then a screening test for gestational diabetes.

Examination

N/A

Laboratory Investigations

Episode No	MRN	HPRN	Routine		Visit T
		F 32 y 28/07/1987			<SAZ <SAZ
Fluoride blood;					
F.N.	Alt	RN	Collection	05/03/2020 07:22	
Hos	Vanguard MOU (St Monica's)		Received	05/03/2020 18:14	
Wrd	Ward not stated		Registered	06/03/2020 03:26	
Doc	DR			ePR Detail	[* Curr.]

Test Set	Staff Notes	Test Item	Result	Units	Normal Values	Previous Result 1	Previous Result
GLUF		Fasting	3.9	mmol/L		3.7 19/09/2019 07:54	
		Fasting glucose auto c					
		Fasting glucose DFT	Y			Y 19/09/2019 07:54	
GL120		120 min	4.4	mmol/L		3.4 19/09/2019 07:54	

Other Investigations

Literature search on Google Scholar yielded the following interesting article:

Effectiveness of sodium fluoride as a preservative of glucose in blood.

[A Y Chan](#), [R Swaminathan](#), [C S Cockram](#) *Clinical Chemistry*, Volume 35, Issue 2, 1 February 1989, Pages 315–317, <https://doi.org/10.1093/clinchem/35.2.315>
Published: 01 February 1989

Abstract

How effective is sodium fluoride as a preservative of blood glucose? We compared changes in glucose concentration in fluoride-treated blood specimens with those of heparin-treated specimens. The former declined rapidly during the first hour; thereafter the rate of decrease was slower, and after 4 h the glucose concentration in the blood samples remained stable for up to three days. In contrast, the glucose concentration in the heparin-containing samples declined continuously. During the first hour, however, the rates of decline in the two types of samples were similar. **Evidently sodium fluoride takes effect slowly but effectively in preserving glucose in blood for at least three days.** Its use, however, is unnecessary if the concentration of glucose is to be measured within the first hour after sampling.

Final Diagnosis

This is likely a true result, meaning the patient is normal and does not have impaired glucose tolerance, nor diabetes.

Take Home Message

Blood glucose is stable for 3 days in plasma from NaF tubes, whether being centrifuged in a timely manner or not.

The stability of glucose in specimens is affected by storage temperature, bacterial contamination, and glycolysis. Plasma or serum samples without preservative (NaF) should be separated from cells or clot within half an hour of being drawn. When blood is permitted to clot and to stand uncentrifuged at room temperature, the average decrease in serum glucose is ~7% per hour (0.28 – 0.56 mmol/L/hour), as a result of glycolysis. Glycolysis can be inhibited by collecting the specimen in fluoride tubes (1).

(1) Sacks DB. Carbohydrates. In: Tietz NW, ed. Fundamentals of Clinical Chemistry. 4th ed. Philadelphia: WB Saunders 1996;351-374.